



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

SP Metrology System (Thailand) Co., Ltd.
69/29 Moo 1 T.Klongsi A.Klongluang
Pathumthani, Thailand 12120

Fulfills the requirements of

ISO/IEC 17025:2017

In the fields of

CALIBRATION and DIMENSIONAL MEASUREMENT

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 16 May 2026
Certificate Number: ACT-2050



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

SP Metrology System (Thailand) Co., Ltd.

69/29 Moo 1 T.Klongsi A.Klongluang

Pathumthani, Thailand 12120

Saton Samerphark +662-193-2220

Saton.Samerphark@trescal.com www.spmetrology.co.th

CALIBRATION AND DIMENSIONAL MEASUREMENT

Valid to: **May 16, 2026**

Certificate Number: **ACT-2050**

CALIBRATION

Acoustics and Vibration

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
¹ Sound Level Meter	94 dB 114 dB	0.15 dB 0.15 dB	Comparison to Sound Level Calibrator
¹ Vibration Meter w/Sensor Acceleration	40 Hz (0 to 50) m/s ² (RMS) 160 Hz (0 to 50) m/s ² (RMS) 1 kHz (0 to 20) m/s ² (RMS) 5 kHz (0 to 20) m/s ² (RMS)	1.6 % of reading 1.5 % of reading 1.5 % of reading 2.4 % of reading	Comparison to Vibration Calibrator
Velocity	40 Hz (0 to 50) mm/s (RMS) 160 Hz (0 to 50) mm/s (RMS)	1.5 % of reading 2.7 % of reading	
Displacement	40 Hz (0 to 50) μm (RMS) 160 Hz (0 to 50) μm (RMS)	3.2 % of reading 3.2 % of reading	

Chemical Quantities

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
^{1,2} pH Meter	4.01 pH 7.01 pH 10.01 pH	0.012 pH 0.012 pH 0.012 pH	Direct Measure using Accredited pH Solution Buffer
^{1,2} Conductivity Meter	84 µS/cm 1 413 µS/cm 12 880 µS/cm	0.6 µS/cm 8.2 µS/cm 75 µS/cm	Direct Measure using Accredited Conductivity Solution
¹ Refractometer	5 %Brix 10 %Brix 20 %Brix 30 %Brix 60 %Brix	0.075 %Brix 0.074 %Brix 0.074 %Brix 0.073 %Brix 0.072 %Brix	Direct Measure using Sucrose Standard Solution
¹ Refractometer Refractive Index	1.340 27 nD 1.347 84 nD 1.363 85 nD 1.381 14 nD 1.441 89 nD	0.000 11 nD 0.000 11 nD 0.000 11 nD 0.000 11 nD 0.000 11 nD	Direct Measure using Sucrose Standard Solution
¹ Turbidity meter	20 NTU 100 NTU 200 NTU 800 NTU	0.12 NTU 0.4 NTU 1.2 NTU 6 NTU	Direct Measure using Turbidity Standard Solution
¹ Viscometer Rotational @ 25°C	101.1 cP 6 618 cP 15 608 cP	0.16 cP 15 cP 36 cP	Direct Measure using STD Viscosity Solution
¹ Total Dissolved Solids (TDS) Meter	1 000 mg/l	32 mg/l	Direct Measure using TDS Solution
Breath Alcohol Tester/Analyzer	44 mg/dL 70 mg/dL	1 % of reading 1 % of reading	Direct Measure using Alcohol Standard Solutions
¹ Gas Detectors Carbon Monoxide	100 µmol/mol	1 % of reading	Direct Measure using Accredited Gas Mixtures
Methane	22 mmol/mol	1.5 % of reading	
Oxygen	18 cmol/mol	1.1 % of reading	
¹ Salinity Meter Nominal	50 000 mg/l (5 %Salinity) 180 000 mg/l (18 %Salinity)	0.014 %Salinity 0.054 % Salinity	Direct Measure using Accredited Sodium Chloride Solution

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
¹ RTD Temperature Indicator (Simulator) Temperature Indicator	RTD (Pt100) (-200 to 800) °C 10 Ω, Cu 427 (-100 to 260) °C	0.25 °C 0.42 °C	Comparison to Fluke 744 Process Calibrator
¹ Thermocouple Temperature Indicator (Simulator) Temperature Indicator	Type E (-250 to 1 000) °C Type J (-210 to 1 200) °C Type K (-200 to 1 372) °C Type R and S (0 to 1 768) °C Type T (-250 to 400) °C	0.42 °C 0.45 °C 0.38 °C 0.96 °C 0.64 °C	Comparison to Fluke 744 Process Calibrator
¹ DC Voltage Source	Up to < 330 mV 330 mV to < 3.3 V (3.3 to < 33) V (33 to < 330) V (330 to 1 000) V	48 μV/V + 9 μV 40 μV/V + 60 μV 40 μV/V + 0.6 mV 45 μV/V + 6 mV 45 μV/V + 60 mV	Comparison to Fluke 5502A Multiproduct Calibrator
¹ DC Current Source	Up to < 3.3 mA (3.3 to < 33) mA (33 to < 330) mA (0.33 to < 1.1) A (1.1 to < 3) A (3.0 to < 11) A (11 to 20) A	0.08 mA/A + 85 nA 0.08 mA/A + 0.65 μA 0.08 mA/A + 7.8 μA 0.3 mA/A + 80 μA 0.3 mA/A + 85 μA 0.47 mA/A + 0.8 mA 0.78 mA/A + 5.8 mA	Comparison to Fluke 5502A Multiproduct Calibrator
¹ AC Voltage Source	(1 to < 33) mV (10 to 45) Hz > 45 Hz to 10 kHz (> 10 to 20) kHz (> 20 to 50) kHz (> 50 to 100) kHz (> 100 to 450) kHz (33 to < 330) mV (10 to 45) Hz > 45 Hz to 10 kHz (> 10 to 20) kHz (> 20 to 50) kHz (> 50 to 100) kHz (> 100 to 450) kHz	1.5 mV/V + 20 μV 0.8 mV/V + 20 μV 1.5 mV/V + 20 μV 1.6 mV/V + 20 μV 3 mV/V + 30 μV 8 mV/V + 50 μV 0.5 mV/V + 20 μV 0.25 mV/V + 20 μV 0.6 mV/V + 20 μV 0.8 mV/V + 35 μV 1.9 mV/V + 0.15 mV 4 mV/V + 0.3 mV	Comparison to Fluke 5502A Multiproduct Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
¹ AC Voltage Source	(0.33 to < 3.3) V		Comparison to Fluke 5502A Multiproduct Calibrator
	(10 to 45) Hz	0.4 mV/V + 80 μ V	
	> 45 Hz to 10 kHz	0.25 mV/V + 0.1 mV	
	(> 10 to 20) kHz	0.6 mV/V + 0.1 mV	
	(> 20 kHz to 50) kHz	0.8 mV/V + 0.1 mV	
	(> 50 kHz to 100) kHz	1.9 mV/V + 0.2 mV	
	(> 100 kHz to 450) kHz	4 mV/V + 0.8 mV	
	(3.3 to < 33) V		
	(10 Hz to 45) Hz	0.4 mV/V + 0.9 mV	
	> 45 Hz to 10 kHz	0.25 mV/V + 0.8 mV	
	(> 10 kHz to 20) kHz	0.6 mV/V + 0.9 mV	
	(> 20 kHz to 50) kHz	0.8 mV/V + 0.9 mV	
	(> 50 kHz to 90) kHz	1.9 mV/V + 2 mV	
	(33 to < 330) V		
	45 Hz to 1 kHz	0.4 mV/V + 7 mV	
	(> 1 to 10) kHz	0.65 mV/V + 10 mV	
	(> 10 kHz to 18) kHz	0.7 mV/V + 15 mV	
¹ AC Current Source	(330 to 1 000) V		Comparison to Fluke 5502A Multiproduct Calibrator
	45 Hz to 1 kHz	0.4 mV/V + 60 mV	
	(> 1 kHz to 5) kHz	0.65 mV/V + 78 mV	
	(> 5 kHz to 10) kHz	0.7 mV/V + 78 mV	
	(0.029 to < 0.33) mA		
	(20 Hz to 45) Hz	1.6 mA/A + 80 nA	
	> 45 Hz to 1 kHz	1 mA/A + 80 nA	
	(> 1 to 5) kHz	2.4 mA/A + 0.12 μ A	
	(> 5 to 10) kHz	6.3 mA/A + 0.16 μ A	
	(0.33 to < 3.3) mA		
	(20 Hz to 45) Hz	1.6 mA/A + 0.15 μ A	
	> 45 Hz to 1 kHz	0.8 mA/A + 0.15 μ A	
	(> 1 to 5) kHz	1.6 mA/A + 0.2 μ A	
	(> 5 to 10) kHz	4 mA/A + 0.25 μ A	
	(3.3 to < 33) mA		
	(20 Hz to 45) Hz	1.4 mA/A + 2 μ A	
	> 45 Hz to 1) kHz	0.32 mA/A + 1.8 μ A	
	(> 1 to 5) kHz	0.65 mA/A + 1.8 μ A	
	(> 5 to 10) kHz	1.6 mA/A + 2.5 μ A	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
¹ AC Current Source	(33 to < 330) mA	1.5 mA/A + 18 μ A	Comparison to Fluke 5502A Multiproduct Calibrator
	(20 Hz to 45) Hz	0.32 mA/A + 18 μ A	
	> 45 Hz to 1 kHz	0.8 mA/A + 40 μ A	
	(> 1 to 5) kHz	1.6 mA/A + 80 μ A	
	(> 5 to 10) kHz		
	(0.33 to < 1.1) A	1.4 mA/A + 0.15 mA	
	(20 to 45) Hz	0.4 mA/A + 0.15 mA	
	> 45 Hz to 1 kHz	5 mA/A + 0.8 mA	
	(> 1 to 5) kHz		
	(1.1 to < 3) A	1.5 mA/A + 0.15 mA	
¹ Resistance Source	(45 to 65) Hz	0.5 mA/A + 0.15 mA	Comparison to Fluke 5502A Multiproduct Calibrator
	(> 65 to 500) Hz	0.5 mA/A + 0.15 mA	
	> 500 Hz to 1 kHz	0.5 mA/A + 0.15 mA	
	(> 1 to 5) kHz	4.8 mA/A + 6 mA	
	(3 to < 11) A	0.5 mA/A + 1.8 mA	
	(45 to 65) Hz	0.8 mA/A + 2 mA	
	> 65 Hz to 1 kHz		
	(11 to 20) A	0.95 mA/A + 7 mA	
	(45 to 65) Hz	1.2 mA/A + 8.2 mA	
	> 65 Hz to 1 kHz		
	(0 to 11) Ω	0.1 m Ω / Ω + 10 m Ω	
	(11 to 33) Ω	0.1 m Ω / Ω + 15 m Ω	
	(33 to 110) Ω	80 $\mu\Omega$ / Ω + 15 m Ω	
	(110 to 330) Ω	80 $\mu\Omega$ / Ω + 20 m Ω	
	330 Ω to 1.1 k Ω	80 $\mu\Omega$ / Ω + 0.1 Ω	
	(1.1 to 3.3) k Ω	80 $\mu\Omega$ / Ω + 0.2 Ω	
	(3.3 to 11) k Ω	80 $\mu\Omega$ / Ω + 0.8 Ω	
	(11 to 33) k Ω	80 $\mu\Omega$ / Ω + 1.2 Ω	
	(33 to 110) k Ω	0.1 m Ω / Ω + 6 Ω	
	(110 to 330) k Ω	0.11 m Ω / Ω + 12 Ω	
	330 k Ω to 1.1 M Ω	0.13 m Ω / Ω + 70 Ω	
	(1.1 to 3.3) M Ω	0.13 m Ω / Ω + 0.15 k Ω	
	(3.3 to 11) M Ω	0.48 m Ω / Ω + 1 k Ω	
	(11 to 33) M Ω	0.8 m Ω / Ω + 3 k Ω	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
¹ Capacitance Source			Comparison to Fluke 5502A Multiproduct Calibrator
10 Hz to 10 kHz	(0.1 to 0.5) nF	0.4 % of reading + 8pF	
10 Hz to 10 kHz	(0.5 to 1.09) nF	0.4 % of reading + 10 pF	
10 Hz to 3 kHz	(1.1 to 3.29) nF	0.4 % of reading + 10 pF	
10 Hz to 1 kHz	(3.3 to 10.9) nF	0.2 % of reading + 12 pF	
10 Hz to 1 kHz	(11 to 32.9) nF	0.2 % of reading + 0.1 nF	
10 Hz to 1 kHz	(33 to 109.9) nF	0.2 % of reading + 0.1 nF	
10 Hz to 1 kHz	(110 to 329.9) nF	0.2 % of reading + 0.7 nF	
10 Hz to 600 Hz	(0.33 to 1.09) μ F	0.2 % of reading + 1.3 nF	
10 Hz to 300 Hz	(1.1 to 3.29) μ F	0.2 % of reading + 7 nF	
10 Hz to 150 Hz	(3.3 to 10.9) μ F	0.2 % of reading + 10 nF	
10 Hz to 120 Hz	(11 to 32.9) μ F	0.32 % of reading + 0.08 μ F	
10 Hz to 80 Hz	(33 to 109.9) μ F	0.37 % of reading + 0.11 μ F	
50 Hz	(110 to 329.9) μ F	0.37 % of reading + 0.7 μ F	
20 Hz	(0.33 to 1.09) mF	0.37 % of reading + 1 μ F	
6 Hz	(1.1 to 3.29) mF	0.37 % of reading + 6.5 μ F	
2 Hz	(3.3 to 10.9) mF	0.37 % of reading + 10 μ F	
0.6 Hz	(11 to 32.9) mF	0.6 % of reading + 63 μ F	
0.2 Hz	(33 to 50) mF	0.85 % of reading + 98 μ F	
¹ DC Current Clamp Source	Up to 200 A (> 200 to 550) A (> 550 to 1 000) A	3.8 mA/A + 60 mA 3.3 mA/A + 80 mA 3.2 mA/A + 70 mA	Comparison to Fluke 5502A Multiproduct Calibrator with Current Coil
¹ AC Current Clamp Source	Up to 20 A (45 to 100) Hz (> 100 to 440) Hz (> 20 to 200) A (45 to 65) Hz (> 65 to 100) Hz (> 100 to 440) Hz (> 200 to 550) A (45 to 65) Hz (> 65 to 100) Hz (> 550 to 1 000) A (45 to 65) Hz (> 65 to 100) Hz	5 mA/A + 80 mA 12 mA/A + 80 mA 5 mA/A + 80 mA 9.5 mA/A + 80 mA 14 mA/A + 80 mA 3.7 mA/A + 72 mA 9 mA/A + 60 mA 3.7 mA/A + 80 mA 9.3 mA/A + 80 mA	Comparison to Fluke 5502A Multiproduct Calibrator with Current Coil

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
¹ Insulation Resistance Source Test Voltage @ 50 V, 100 V, 250 V, 500 V, 1 000 V	(0.1 to 10) MΩ (10 to 20) MΩ (20 to 30) MΩ (30 to 50) MΩ (50 to 100) MΩ (100 to 200) MΩ (200 to 500) MΩ (500 to 1 000) MΩ	5.8 kΩ 32 kΩ 69 kΩ 75 kΩ 94 kΩ 2.8 MΩ 4.5 MΩ 7.1 MΩ	Comparison to Resistance Decade Box
¹ Electrical Simulation of Thermocouple Indicating Devices – Source/Measure	Type K (-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 1 000) °C (1 000 to 1 372) °C Type B (600 to 800) °C (800 to 1 000) °C (1 000 to 1 550) °C (1 550 to 1 820) °C Type E (-250 to -100) °C (-100 to -25) °C (-25 to 350) °C (350 to 650) °C (650 to 1 000) °C Type J (-210 to -100) °C (-100 to -25) °C (-25 to 150) °C (150 to 760) °C (760 to 1 200) °C Type N (-250 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 410) °C (410 to 1 300) °C	0.46 °C 0.26 °C 0.23 °C 0.37 °C 0.56 °C 0.61 °C 0.48 °C 0.42 °C 0.46 °C 0.7 °C 0.23 °C 0.2 °C 0.23 °C 0.3 °C 0.56 °C 0.32 °C 0.28 °C 0.24 °C 0.33 °C 0.56 °C 0.32 °C 0.28 °C 0.26 °C 0.38 °C	Comparison to Fluke 5502A Multiproduct Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
¹ Electrical Simulation of Thermocouple Indicating Devices – Source/Measure	Type R		Comparison to Fluke 5502A Multiproduct Calibrator
	(0 to 250) °C	0.8 °C	
	(250 to 400) °C	0.5 °C	
	(400 to 1 000) °C	0.47 °C	
	(1 000 to 1 767) °C	0.56 °C	
	Type S		
	(0 to 250) °C	0.66 °C	
	(250 to 400) °C	0.51 °C	
	(400 to 1 000) °C	0.52 °C	
	(1 000 to 1 767) °C	0.65 °C	
	Type U		
	(-200 to 0) °C	0.79 °C	
	(0 to 600) °C	0.38 °C	
	Type T		
	(-250 to -150) °C	0.88 °C	
	(-150 to 0) °C	0.34 °C	
	(0 to 120) °C	0.23 °C	
	(120 to 400) °C	0.21 °C	
¹ Electrical Simulation of RTD Indicating Devices – Source	100 Ω, Pt385		Comparison to Fluke 5502A Multiproduct Calibrator
	(-200 to -80) °C	0.09 °C	
	(-80 to 0) °C	0.09 °C	
	(0 to 100) °C	0.14 °C	
	(100 to 300) °C	0.14 °C	
	(300 to 400) °C	0.15 °C	
	(400 to 630) °C	0.18 °C	
	(630 to 800) °C	0.33 °C	
	200 Ω, Pt385		
	(-200 to -80) °C	0.08 °C	
	(-80 to 0) °C	0.08 °C	
	(0 to 100) °C	0.08 °C	
	(100 to 260) °C	0.09 °C	
	(260 to 300) °C	0.18 °C	
	(300 to 400) °C	0.19 °C	
	(400 to 600) °C	0.2 °C	
	(600 to 630) °C	0.23 °C	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
¹ Electrical Simulation of RTD Indicating Devices – Source	500 Ω , Pt385		Comparison to Fluke 5502A Multiproduct Calibrator
	(-200 to -80) °C	0.08 °C	
	(-80 to 0) °C	0.09 °C	
	(0 to 100) °C	0.09 °C	
	(100 to 260) °C	0.1 °C	
	(260 to 300) °C	0.13 °C	
	(300 to 400) °C	0.13 °C	
	(400 to 600) °C	0.14 °C	
	(600 to 630) °C	0.16 °C	
	1 000 Ω , Pt385		
	(-200 to -80) °C	0.08 °C	
	(-80 to 0) °C	0.09 °C	
	(0 to 100) °C	0.09 °C	
	(100 to 260) °C	0.1 °C	
	(260 to 300) °C	0.13 °C	
	(300 to 400) °C	0.13 °C	
	(400 to 600) °C	0.14 °C	
	(600 to 630) °C	0.16 °C	
	100 Ω , Pt3916		
	(-200 to -190) °C	0.36 °C	
	(-190 to -80) °C	0.08 °C	
	(- 80 to 0) °C	0.09 °C	
	(0 to 100) °C	0.11 °C	
	(100 to 260) °C	0.11 °C	
	(260 to 300) °C	0.13 °C	
	(300 to 400) °C	0.14 °C	
	(400 to 600) °C	0.15 °C	
	(600 to 630) °C	0.33 °C	
	100 Ω , Pt3926		
	(-200 to - 80) °C	0.08 °C	
	(-80 to 0) °C	0.09 °C	
	(0 to 100) °C	0.12 °C	
	(100 to 300) °C	0.14 °C	
	(300 to 400) °C	0.16 °C	
	(400 to 630) °C	0.18 °C	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
¹ DC Power Source	(0.33 to 1 000) V, Up to 0.33 A (0.108 9 to < 330) W (0.33 to 1 000) V, (0.33 to < 3) A 330 W to < 3 kW (0.33 to 1 000) V, (3 to < 10.9) A (3.0 to < 10.9) kW (0.33 to 1 000) V, (10.99 to 20) A (10.9 to 20) kW	0.10 mW/W + 6 mW 0.31 mW/W + 60 mW 0.5 mW/W + 0.6 W 0.81 mW/W + 0.6 W	Comparison to Fluke 5502A Multiproduct Calibrator
¹ AC Power Source	(45 to 65) Hz, PF=1 (0.33 to 1 000) V Up to 0.329 A (0.109 to < 10.9) W (10.9 to < 330) W (0.33 to 1 000) V, (0.33 to < 3) A 330 W to < 3 kW (0.33 to 1 000) V, (3 to < 10.9) A (3 to < 3.5) kW (3.5 to < 10.9) kW (0.33 to 1 000) V, (10.9 to 20) A (10.9 to 20) kW	0.55 mW/W + 1.5 mW 0.8 mW/W + 6 mW 1.2 mW/W + 60 mW 1.4 mW/W + 60 mW 1.4 mW/W + 0.6 W 1.2 mW/W + 0.6 W	Comparison to Fluke 5502A Multiproduct Calibrator
¹ DC Voltage Measure	Up to 100 mV (> 0.1 to 1) V (> 1 to 10) V (> 10 to 100) V (> 100 to 1 000) V	14 μ V/V + 0.35 μ V 4.9 μ V/V + 0.35 μ V 4.7 μ V/V + 0.58 μ V 7 μ V/V + 35 μ V 21 μ V/V + 0.12 mV	Comparison to HP 3458A 8.5 Digit Multimeter
¹ AC Voltage Measure	100 mV to 10 V (10 to 50) Hz > 50 Hz to 1 kHz (> 1 to 20) kHz (> 20 to 50) kHz (> 50 to 100) kHz (> 100 to 300) kHz > 300 kHz to 1 MHz	85 μ V/V + 0.46 mV 85 μ V/V + 0.23 mV 0.16 mV/V + 0.23 mV 0.35 mV/V + 0.23 mV 0.93 mV/V + 0.23 mV 3.5 mV/V + 1.2 mV 12 mV/V + 1.2 mV	Comparison to HP 3458A 8.5 Digit Multimeter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
¹ AC Voltage Measure	(> 10 to 100) V (10 to 50) Hz > 50 Hz to 1 kHz (> 1 to 20) kHz (> 20 to 50) kHz (> 50 to 100) kHz (> 100 to 1 000) V (10 to 50) Hz > 50 Hz to 1 kHz (> 1 to 20) kHz (> 20 to 50) kHz (> 50 to 100) kHz	0.23 mV/V + 4.6 mV 0.23 mV/V + 2.3 mV 0.23 mV/V + 2.3 mV 0.41 mV/V + 2.3 mV 1.4 mV/V + 2.3 mV 0.46 mV/V + 46 mV 0.46 mV/V + 23 mV 0.69 mV/V + 23 mV 1.4 mV/V + 23 mV 3.5 mV/V + 23 mV	Comparison to HP 3458A 8.5 Digit Multimeter
¹ DC Current Measure	(> 10 to 100) μ A (> 0.1 to 1) mA (> 1.0 to 10) mA (> 10 to 100) mA (> 0.10 to 1.0) A	24 μ A/A + 0.92 nA 24 μ A/A + 5.8 nA 24 μ A/A + 58 nA 41 μ A/A + 0.58 μ A 0.13 mA/A + 12 μ A	Comparison to HP 3458A 8.5 Digit Multimeter
¹ DC Current Measure	(>1 to 3) A	2.4 mA/A + 1.1 mA	Comparison to Keysight 34460A 6.5 Digit Multimeter
¹ AC Current Measure	(0.1 to 1) mA (10 to 20) Hz (>20 to 45) Hz (>45 to 100) Hz >100 Hz to 1 kHz (1 to 100) mA (10 to 20) Hz (>20 to 45) Hz (>45 to 100) Hz >100 Hz to 1 kHz (0.1 to 1) A (10 to 20) Hz (>20 to 45) Hz (>45 to 100) Hz >100 Hz to 1 kHz	4.6 mA/A + 0.35 μ A 1.7 mA/A + 0.35 μ A 0.7 mA/A + 0.35 μ A 0.7 mA/A + 0.35 μ A 4.6 mA/A + 0.23 mA 1.7 mA/A + 23 μ A 0.7 mA/A + 23 μ A 0.35 mA/A + 23 μ A 4.6 mA/A + 0.23 mA 1.8 mA/A + 0.23 mA 0.93 mA/A + 0.23 mA 1.2 mA/A + 0.23 mA	Comparison to HP 3458A 8.5 Digit Multimeter
¹ AC Current Measure	(>1 to 3) A 50 Hz to 5 kHz	2.7 mA/A + 2 mA	Comparison to Keysight 34460A 6.5 Digit Multimeter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
¹ Resistance Measure	Up to 10 Ω (> 10 to 100) Ω > 100 Ω to 1 k Ω (> 1 to 10) k Ω (> 10 to 100) k Ω (> 0.1 to 1) M Ω (> 1 to 10) M Ω (> 10 to 100) M Ω > 100 M Ω to 1 G Ω	18 $\mu\Omega/\Omega$ + 58 $\mu\Omega$ 14 $\mu\Omega/\Omega$ + 0.58 m Ω 12 $\mu\Omega/\Omega$ + 0.58 m Ω 1 $\mu\Omega/\Omega$ + 5.8 m Ω 12 $\mu\Omega/\Omega$ + 58 m Ω 19 $\mu\Omega/\Omega$ + 2.3 Ω 59 $\mu\Omega/\Omega$ + 0.12 k Ω 0.58 m Ω/Ω + 1.2 k Ω 5.8 m Ω/Ω + 12 k Ω	Comparison to HP 3458A 8.5 Digit Multimeter
¹ DC High Voltage Measure	Up to 1 kV (> 1 to 3) kV (> 3 to 5) kV (> 5 to 10) kV	24 mV/V + 60 μ V 24 mV/V + 70 μ V 24 mV/V + 90 μ V 24 mV/V + 0.18 mV	Comparison to Keysight 34460A 6.5 Digit Multimeter with Fluke 80K-40 High Voltage Probe
¹ AC High Voltage Measure	(1 to 6) kV (50 to 60) Hz	58 mV/V + 4 mV	Comparison to Keysight 34460A 6.5 Digit Multimeter with Fluke 80K-40 High Voltage Probe
LCR Meter Inductance (L) (Variable Artifact)	100 μ H to 10 H @ 100 mV, 1 V, 1 kHz	1.2 % of reading + 0.9 μ H	Comparison to IET 1492 Decade Inductor
LCR Meter Capacitance (C) (Variable Artifact)	1 pF to 1 μ F @ 1 V, 1 kHz	0.06 % of reading + 0.6 pF	Comparison to General Radio 1413 Precision Decade Capacitor
LCR Meter Resistance (R) (Variable Artifact)	1 Ω to 100 k Ω @ 1 V, 1 kHz	0.2 % of reading + 2.4 m Ω	Comparison to Decade Resistance Box
Inductance Measure	100 μ H to 10 H @ 1 V, 1 kHz	0.06 % of reading	Comparison to Agilent E4980A LCR Meter
Capacitance Measure	1pF to 1 μ F @ 1 V, 100 Hz to 1 MHz	0.06 % of reading	Comparison to Agilent E4980A LCR Meter
¹ Oscilloscope Vertical Deflection DC 50 Ω and 1 M Ω	2 mV 5 mV 10 mV 20 mV 50 mV	0.84 % of reading 0.49 % of reading 0.34 % of reading 0.26 % of reading 0.29 % of reading	Comparison to Fluke 5502A Multiproduct Calibrator, Fluke PM6685R Universal Frequency Counter, HP 3458A 8.5 Digit Multimeter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
¹ Oscilloscope Vertical Deflection DC 50 Ω and 1 M Ω	100 mV 200 mV 500 mV 1 V 2 V 5 V 10 V 20 V	0.23 % of reading 0.21 % of reading 0.28 % of reading 0.27 % of reading 0.22 % of reading 0.28 % of reading 0.2 % of reading 0.2 % of reading	Comparison to Fluke 5502A Multiproduct Calibrator, Fluke PM6685R Universal Frequency Counter, HP 3458A 8.5 Digit Multimeter
¹ Oscilloscope Vertical Bandwidth 3 dB down from Reference Amplitude	50 kHz to 100 MHz (> 100 to 300) MHz	1.4 % of reading 1.8 % of reading	Comparison to Fluke 5502A Multiproduct Calibrator,
¹ Oscilloscope Horizontal Deflection: Time Mark	1 ns 2 ns 5 ns 10 ns 20 ns 50 ns 100 ns 200 ns 500 ns 1 μ s 2 μ s 5 μ s	0.01 % of reading 0.04 % of reading 0.01 % of reading 0.01 % of reading 0.04 % of reading 0.01 % of reading 0.01 % of reading 0.04 % of reading 0.01 % of reading 0.01 % of reading 0.04 % of reading 0.01 % of reading	Comparison to Fluke 5502A Multiproduct Calibrator
¹ Oscilloscope Horizontal Deflection: Time Mark	10 μ s 20 μ s 50 μ s 100 μ s 200 μ s 500 μ s 1 ms 2 ms 5 ms 10 ms 20 ms 50 ms 100 ms 200 ms 500 ms	0.01 % of reading 0.04 % of reading 0.01 % of reading 0.01 % of reading 0.04 % of reading 0.01 % of reading 0.01 % of reading 0.04 % of reading 0.01 % of reading 0.01 % of reading 0.04 % of reading 0.01 % of reading 0.01 % of reading 0.04 % of reading 0.01 % of reading	Comparison to Fluke 5502A Multiproduct Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
¹ Oscilloscope Horizontal Deflection: Time Mark	1 s 2 s 5 s 10 s	0.62 % of reading 1.2 % of reading 3.1 % of reading 6.2 % of reading	Comparison to Fluke 5502A Multiproduct Calibrator
¹ Oscilloscope DC Accuracy 50 Ω and 1 M Ω (Digital)	2 mV 5 mV 10 mV 20 mV 50 mV 100 mV 200 mV 500 mV 1 V 2 V 5 V 10 V 20 V	0.11 % of reading 0.39 % of reading 0.19 % of reading 0.1 % of reading 0.38 % of reading 0.19 % of reading 0.1 % of reading 0.38 % of reading 0.19 % of reading 0.1 % of reading 0.04 % of reading 0.02 % of reading 0.01 % of reading	Comparison to Fluke 5502A Multiproduct Calibrator
^{1,3} Oscilloscope Time Base	10 MHz	$2.4 \times 10^{-10} f$	Comparison to Fluke PM6685R Universal Frequency Counter
¹ Oscilloscope Probe Compensator Output	≤ 10 Vp-p @ ≤ 10 kHz	0.02 % of reading	Comparison to HP 3458A 8.5 Digit Multimeter
Single and Three Phase Power Analyzer	AC Voltage @ 50/60 Hz Up to 1 000 V AC Current @ 50/60 Hz Up to 400 A AC Power @ 50/60 Hz Up to 271.6 kW Power Factor (0.5 to 1)	0.58 % of reading 1.3 % of reading 1.3 % of reading 1.2 % of reading	Comparison to Three Phase Energy and Power Analyzer with AC Current Clamp (Compare with Power Meter Standard)

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
⁴ Tuned RF Level Measure 10 dB/step	(-40 to 0) dB		Comparison to HP 8902A Measuring Receiver, HP 11722A/11792A Power Sensor
	10 MHz	0.11 dB	
	50 MHz	0.10 dB	
	100 MHz	0.10 dB	
	400 MHz	0.10 dB	
	1 000 MHz	0.10 dB	
	2 000 MHz	0.11 dB	
	3 000 MHz	0.11 dB	
	4 000 MHz	0.13 dB	
	5 000 MHz	0.11 dB	
	6 000 MHz	0.13 dB	
	7 000 MHz	0.17 dB	
	8 000 MHz	0.15 dB	
	9 000 MHz	0.12 dB	
⁴ Tuned RF Level Measure 10 dB/step	(-40 to 0) dB		Comparison to HP 8902A Measuring Receiver, HP 11722A/11792A Power Sensor
	1 0000 MHz	0.12 dB	
	11 000 MHz	0.11 dB	
	12 000 MHz	0.16 dB	
	13 000 MHz	0.15 dB	
	14 000 MHz	0.14 dB	
	15 000 MHz	0.23 dB	
	16 000 MHz	0.26 dB	
	17 000 MHz	0.17 dB	
	18 000 MHz	0.21 dB	
Distortion Source	@ 20 Hz to 20 kHz (-80 to -40) dB	1.2 dB	Comparison to HP 8903B Audio Analyzer
	@ (>20 to 100) kHz (-80 to -40) dB	2.3 dB	

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
¹ External Micrometer V-Anvil, Screw Thread, Indicating	Up to 2.5 mm (2.5 to 5.1) mm (5.1 to 7.7) mm (7.7 to 10.3) mm (10.3 to 12.9) mm (12.9 to 15) mm (15 to 17.6) mm (17.6 to 20.2) mm (20.2 to 22.8) mm (22.8 to 25) mm (25 to 50) mm (50 to 75) mm (75 to 100) mm (100 to 125) mm (125 to 150) mm (150 to 175) mm (175 to 200) mm (200 to 250) mm (250 to 300) mm (300 to 400) mm (400 to 500) mm (500 to 600) mm (600 to 1 000) mm	94 nm 0.12 µm 0.14 µm 0.17 µm 0.2 µm 0.23 µm 0.26 µm 0.29 µm 0.32 µm 0.35 µm 0.89 µm 1.2 µm 1.5 µm 1.8 µm 2.1 µm 2.4 µm 2.8 µm 3.4 µm 4.1 µm 5.4 µm 6.8 µm 8.1 µm 13.5 µm	Comparison to Gauge Block Set, Optical Flats
¹ Vernier Caliper Dial and Digital	Up to 200 mm (200 to 300) mm (300 to 400) mm (400 to 500) mm (500 to 600) mm (600 to 700) mm (700 to 800) mm (800 to 900) mm (900 to 1 000) mm (1 000 to 1 500) mm	6 µm 7 µm 8 µm 9 µm 10 µm 11 µm 12 µm 13 µm 15 µm 21 µm	Comparison to Gauge Block Set
¹ Can Seam Micrometer	Up to 13 mm	2.3 µm	Comparison to Gauge Block Set

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
¹ Internal Micrometer All type Snap Micrometer (Up to 100 mm)	Up to 30 mm (30 to 45) mm (45 to 50) mm (50 to 60) mm (60 to 70) mm (70 to 80) mm (80 to 87) mm (87 to 97) mm (97 to 100) mm (100 to 125) mm (125 to 150) mm (150 to 175) mm (175 to 200) mm (200 to 250) mm (250 to 300) mm (300 to 400) mm (400 to 500) mm (500 to 600) mm	0.7 µm 0.8 µm 0.9 µm 1 µm 1.1 µm 1.2 µm 1.3 µm 1.4 µm 1.5 µm 1.8 µm 2.1 µm 2.4 µm 2.8 µm 3.4 µm 4.1 µm 5.4 µm 6.8 µm 8.1 µm	Comparison to Gauge Block Set
¹ Caliper Gauge External	Up to 25 mm (25 to 50) mm	1.2 µm 1.3 µm	Comparison to Gauge Blocks
¹ Caliper Gauge Internal (0.005 mm) Internal (0.01 mm)	2.5 to 15 mm (10 to 180) mm	3 µm 6 µm	Comparison to Gauge Blocks
¹ Thickness Gauge	Up to 20 mm (20 to 25) mm	0.6 µm 0.7 µm	Comparison to Gauge Blocks
¹ Height Gauge (Dial and Digital)	Up to 20 mm (20 to 50) mm (50 to 100) mm (100 to 150) mm (150 to 200) mm (200 to 250) mm (250 to 300) mm (300 to 400) mm (400 to 500) mm (500 to 600) mm (600 to 700) mm (700 to 800) mm (800 to 900) mm (900 to 1 000) mm	0.6 µm 0.9 µm 1.5 µm 2.1 µm 2.8 µm 3.4 µm 4.1 µm 5.4 µm 6.8 µm 8.1 µm 9.5 µm 11 µm 12 µm 14 µm	Comparison to Gauge Block Set
Feeler Gauge / Thickness Plate	Up to 1 mm (1 to 5) mm	0.21 µm 0.22 µm	Direct Measure using ULM

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Measuring Foil Standard Foil	Up to 1 mm (1 to 5) mm	0.21 μ m 0.22 μ m	Direct Measure using ULM
¹ Indicator	Up to 20 mm (20 to 30) mm (30 to 40) mm (40 to 50) mm (50 to 60) mm (60 to 70) mm (70 to 80) mm (80 to 90) mm (90 to 100) mm	0.6 μ m 0.7 μ m 0.8 μ m 0.9 μ m 1 μ m 1.1 μ m 1.2 μ m 1.3 μ m 1.5 μ m	Comparison to Gauge Blocks
¹ Linear Length Gauge / Electrical Comparators / Mu Checker	Up to 5 mm (5 to 12) mm (12 to 20) mm (20 to 25) mm (25 to 50) mm	0.1 μ m 0.2 μ m 0.3 μ m 0.4 μ m 0.7 μ m	Comparison to Gauge Blocks
Steel Ruler	Up to 100 mm (100 to 200) mm (200 to 300) mm (300 to 400) mm (400 to 500) mm (500 to 600) mm (600 to 700) mm (700 to 800) mm (800 to 900) mm (900 to 1 200) mm (1 200 to 1 500) mm (1 500 to 1 800) mm (1 800 to 2 000) mm	3 μ m 4 μ m 5 μ m 6 μ m 7 μ m 9 μ m 10 μ m 11 μ m 12 μ m 16 μ m 20 μ m 24 μ m 27 μ m	Direct Measure using 3D Vision Measuring Machine
Steel Tape & Textile Tape	Up to 200 mm (200 to 400) mm (400 to 600) mm (600 to 800) mm (800 to 1 000) mm (1 000 to 1 200) mm (1 200 to 1 400) mm	8 μ m 9 μ m 11 μ m 13 μ m 16 μ m 18 μ m 20 μ m	Direct Measure using 3D Vision Measuring Machine

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Steel Tape & Textile Tape	(1 400 to 1 600) mm (1 600 to 1 800) mm (1 800 to 2 000) mm (2 000 to 3 000) mm (3 000 to 4 000) mm (4 000 to 5 000) mm (5 000 to 6 000) mm (6 000 to 7 000) mm (7 000 to 8 000) mm (8 000 to 9 000) mm (9 000 to 10 000) mm (10 000 to 15 000) mm (15 000 to 20 000) mm (20 000 to 25 000) mm (25 000 to 30 000) mm (30 000 to 35 000) mm (35 000 to 40 000) mm (40 000 to 45 000) mm (45 000 to 50 000) mm	23 µm 25 µm 28 µm 41 µm 54 µm 67 µm 80 µm 94 µm 0.11 mm 0.12 mm 0.14 mm 0.2 mm 0.3 mm 0.34 mm 0.4 mm 0.47 mm 0.54 mm 0.6 mm 0.67 mm	Direct Measure using 3D Vision Measuring Machine
¹ Depth Micro Checker, Step Gauge, Inside Checker, Anvil Block	Up to 100 mm (100 to 200) mm (200 to 250) mm (250 to 300) mm Up to 25 mm	2 µm 3 µm 4 µm 5 µm 2 µm	Comparison to Gauge Blocks, Linear Height Master
¹ Depth Gauge, Depth Micrometer	Up to 25 mm (25 to 50) mm (50 to 100) mm (100 to 150) mm (150 to 250) mm (250 to 300) mm (300 to 400) mm (400 to 450) mm	0.7 µm 0.9 µm 1 µm 2 µm 3 µm 4 µm 5 µm 6 µm	Comparison to Gauge Block Set
¹ Surface Plate Overall Flatness Local Area Flatness (Repeat Reading)	Up to 4 m Diagonal (> 4 to 10) m Diagonal Up to 0.1 µm	1.5 µm 6.5 µm 1 µm	In accordance with JIS B 7513 using Planekator (Straight Edge) Dial Indicator

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Plain Plug Gauge, Pin Gauge, Three Wires, T-probe	Up to 15 mm (15 to 22) mm (22 to 30) mm (30 to 40) mm (40 to 50) mm (50 to 60) mm (60 to 70) mm (70 to 80) mm (80 to 90) mm (90 to 100) mm (100 to 150) mm (150 to 200) mm (200 to 250) mm (250 to 300) mm	0.3 μ m 0.4 μ m 0.5 μ m 0.6 μ m 0.7 μ m 0.9 μ m 1 μ m 1.2 μ m 1.3 μ m 1.4 μ m 2 μ m 2.7 μ m 3.4 μ m 4 μ m	Comparison to ULM, Master Gauge Blocks
Plain Ring Gauge	Up to 3 mm (3 to 6) mm (6 to 10) mm (10 to 12) mm (12 to 16) mm (16 to 18) mm (18 to 20) mm (20 to 22) mm (22 to 25) mm (25 to 28) mm (28 to 30) mm (30 to 75) mm (75 to 100) mm (100 to 300) mm	0.44 μ m 0.45 μ m 0.46 μ m 0.48 μ m 0.5 μ m 0.51 μ m 0.53 μ m 0.57 μ m 0.59 μ m 0.61 μ m 0.63 μ m 2.3 μ m 3.1 μ m 4.9 μ m	Comparison to ULM, Master Plain Ring Gauges
¹ Check Master /Caliper Checker	Up to 100 mm (100 to 125) mm (125 to 150) mm (150 to 175) mm (175 to 200) mm (200 to 250) mm (250 to 300) mm (300 to 400) mm (400 to 500) mm (500 to 600) mm (600 to 700) mm	2.7 μ m 2.9 μ m 3.1 μ m 3.3 μ m 3.6 μ m 4.1 μ m 4.7 μ m 5.9 μ m 7.1 μ m 8.4 μ m 9.7 μ m	Comparison to Linear Height Master, Gauge Blocks

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Thread Plug Gauges Pitch Diameter	(M2 to M10) mm	0.47 μ m	Comparison to ULM, Thread Measuring Wire, Gauge Blocks
	(M10 to M20) mm	0.49 μ m	
Major Diameter	(M20 to M50) mm	0.51 μ m	
	(M50 to M70) mm	0.55 μ m	
	(M70 to M100) mm	0.61 μ m	
	(M100 to M150) mm	0.63 μ m	
	(M2 to M10) mm	0.3 μ m	
	(M10 to M20) mm	0.4 μ m	
	(M20 to M30) mm	0.5 μ m	
	(M30 to M40) mm	0.6 μ m	
	(M40 to M50) mm	0.7 μ m	
	(M50 to M60) mm	0.9 μ m	
	(M60 to M70) mm	1 μ m	
	(M70 to M80) mm	1.2 μ m	
	(M80 to M90) mm	1.3 μ m	
	(M90 to M100) mm	1.4 μ m	
	(M100 to M125) mm	1.7 μ m	
	(M125 to M150) mm	2 μ m	
Thread Ring Gauge Pitch Diameter	(M2 to M5) mm	0.57 μ m	Comparison to ULM, Plain Ring Gauges
	(M5 to M8) mm	0.58 μ m	
	(M8 to M10) mm	0.59 μ m	
	(M10 to M12) mm	0.6 μ m	
	(M12 to M18) mm	0.63 μ m	
	(M18 to M20) mm	0.64 μ m	
	(M20 to M25) mm	0.68 μ m	
	(M25 to M30) mm	0.72 μ m	
	(M30 to M75) mm	2.3 μ m	
	(M75 to M90) mm	2.4 μ m	
	(M90 to M100) mm	2.5 μ m	
	(M100 to M125) mm	2.7 μ m	
	(M125 to M150) mm	2.9 μ m	

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Thread Ring Gauge Minor Diameter	(M2 to M8) mm (M8 to M20) mm (M20 to M30) mm (M30 to M75) mm (M75 to M90) mm (M90 to M100) mm (M100 to M125) mm (M125 to M150) mm	0.4 μ m 0.5 μ m 0.6 μ m 1.1 μ m 1.3 μ m 1.4 μ m 1.8 μ m 2.1 μ m	Comparison to ULM, Master Plain Ring Gauges
¹ Dial Gauge Tester, Calibration Tester	Up to 5 mm (5 to 12) mm (12 to 20) mm (20 to 25) mm	0.65 μ m 0.66 μ m 0.68 μ m 0.7 μ m	Comparison to Liner Gauge w/Display
Plain Snap Gauge/Gap Gauge (External)	(2 to 4) mm (4 to 16) mm (16 to 22) mm (22 to 30) mm (30 to 75) mm (75 to 100) mm (100 to 200) mm (200 to 300) mm (300 to 400) mm (400 to 500) mm	0.2 μ m 0.3 μ m 0.4 μ m 0.5 μ m 1.1 μ m 1.4 μ m 2.7 μ m 4 μ m 5.4 μ m 6.7 μ m	Comparison to ULM, Gauge Blocks
Plain Snap Gauge / Gap Gauge (Internal)	(2 to 6) mm (6 to 20) mm (20 to 30) mm (30 to 75) mm (75 to 100) mm (100 to 300) mm	0.4 μ m 0.5 μ m 0.6 μ m 2.3 μ m 2.4 μ m 4.5 μ m	Comparison to ULM, Master Plain Ring Gauges
¹ Hole test, Three-Point Micrometer	(2 to 3) mm (3 to 8) mm (8 to 18) mm (18 to 20) mm (20 to 25) mm (25 to 28) mm (28 to 30) mm 75 mm	0.8 μ m 0.9 μ m 1.2 μ m 1.3 μ m 1.4 μ m 1.5 μ m 1.7 μ m 3 μ m	Direct Measure using Master Ring Gauges
Dial Test Indicator	Up to 1.6 mm	0.3 μ m	Direct Measure using ULM

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
¹ Universal Length Measuring Machine	Up to 1 mm (1 to 3) mm (3 to 5) mm (5 to 10) mm (10 to 25) mm (25 to 50) mm (50 to 75) mm (75 to 100) mm (100 to 125) mm (125 to 150) mm (150 to 175) mm (175 to 200) mm (200 to 250) mm (250 to 300) mm (300 to 400) mm (400 to 500) mm	0.06 μm 0.07 μm 0.09 μm 0.15 μm 0.34 μm 0.67 μm 1 μm 1.3 μm 1.7 μm 2.1 μm 2.4 μm 2.7 μm 3.4 μm 4.1 μm 5.4 μm 6.7 μm	Comparison to Gauge Blocks
¹ Vernier Depth Gauge	Up to 200 mm (200 to 300) mm (300 to 400) mm (400 to 500) mm (500 to 600) mm (600 to 700) mm (700 to 800) mm (800 to 900) mm (900 to 1 000) mm	6 μm 7 μm 8 μm 9 μm 10 μm 11 μm 12 μm 13 μm 15 μm	Comparison to Gauge Block Set
Bore Gauge / Cylinder Gauge	(0.5 to 10) mm (10 to 30) mm (30 to 50) mm (50 to 70) mm (70 to 100) mm (100 to 125) mm (125 to 150) mm (150 to 175) mm (175 to 200) mm (200 to 250) mm (250 to 300) mm (300 to 400) mm (400 to 500) mm (500 to 600) mm (600 to 700) mm (700 to 800) mm	0.6 μm 0.8 μm 0.9 μm 1.1 μm 1.5 μm 1.8 μm 2.1 μm 2.5 μm 2.8 μm 3.5 μm 4.1 μm 5.4 μm 6.8 μm 8.1 μm 9.5 μm 11 μm	Comparison to ULM, Gauge Blocks

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
^{1,3} Profile Projector Linearity	Up to 50 mm (50 to 200) mm (200 to 410) mm	2 µm 3 µm 7 µm	Comparison to Glass Scale
Angle	(0.25 to 30)°	12"	Angle Block Set
¹ Measuring Microscope, Optical Comparator, 3D Vision Measuring System (X, Y)	Up to 50 mm (50 to 200) mm (200 to 410) mm	2 µm 3 µm 7 µm	Comparison to Glass Scale
¹ Height Master	Up to 175 mm (175 to 250) mm (250 to 300) mm (300 to 400) mm (400 to 500) mm (500 to 600) mm (600 to 700) mm	3 µm 4 µm 5 µm 6 µm 7 µm 8 µm 10 µm	Comparison to Gauge Block / Linear Height Master
³ Bevel Protractor	Up to 30° (30 to 45)° (45 to 90)°	12" 24" 48"	Comparison to Angle Block
	Up to 100 mm (100 to 200) mm (200 to 300) mm	3 µm 4 µm 5 µm	Direct Measure using 3D Vision Measuring Machine
Chamfer Gauge	Up to 10 mm	3 µm	Direct Measure using 3D Vision Measuring Machine
Pitch Gauge	Up to 7 mm	3 µm	Direct Measure using 3D Vision Measuring Machine
Radius Gauge	Up to 100 mm	3 µm	Direct Measure using 3D Vision Measuring Machine
Taper Gauge (Scale Type)	Up to 100 mm	3 µm	3D Vision Measuring Machine
Taper Thread Ring	M2 to M5 M5 to M11 M11 to M22 M22 to M45 M45 to M180	0.9 µm 0.91 µm 1.8 µm 5.1 µm 11 µm	Comparison to ULM, Ring Gauge

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
¹ Riser Block	150 mm 300 mm 600 mm	9 µm 10 µm 12 µm	Comparison to Linear Height Master, Gauge Blocks
Long Gauge Block (Grade 1, 2)	100 mm (100 to 125) mm (125 to 150) mm (150 to 175) mm (175 to 200) mm (200 to 250) mm (250 to 300) mm (300 to 400) mm (400 to 500) mm	1.3 µm 1.7 µm 2.1 µm 2.4 µm 2.7 µm 3.4 µm 4.1 µm 5.4 µm 6.7 µm	Comparison to ULM, Master Gauge Blocks
Standard Micrometer, Setting Rod, Length Bar	Up to 25 mm (25 to 50) mm (50 to 75) mm (75 to 100) mm (100 to 125) mm (125 to 150) mm (150 to 175) mm (175 to 200) mm (200 to 300) mm (300 to 400) mm (400 to 500) mm	0.4 µm 0.7 µm 1.1 µm 1.4 µm 1.7 µm 2 µm 2.4 µm 2.7 µm 4 µm 5.4 µm 6.7 µm	Comparison to ULM, Gauge Blocks
Angle Block / Angular	(0.25 to 30)° (30 to 45)° (45 to 60)° (60 to 90)°	12" 24" 36" 48"	Comparison to Angle Block, 3D Vision Measuring Machine
Gauge Blocks	1 mm (1 to 5) mm (5 to 10) mm (10 to 25) mm (25 to 50) mm (50 to 75) mm (75 to 100) mm	0.22 µm 0.23 µm 0.26 µm 0.43 µm 0.72 µm 1.1 µm 1.4 µm	Comparison to ULM, Master Gauge Blocks
Test Sieve	Up to 50 mm	3 µm	Direct Measure using 3D Vision Measuring Machine

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Taper Plug Gauges	Up to M7 M7 to M15 M15 to M25 M25 to M30 M30 to M40 M40 to M50 M50 to M60 M60 to M70 M70 to M80 M80 to M90 M90 to M100 M100 to M200 M200 to M300	0.2 μm 0.3 μm 0.4 μm 0.5 μm 0.6 μm 0.7 μm 0.9 μm 1 μm 1.1 μm 1.3 μm 1.4 μm 2.7 μm 4 μm	Comparison to ULM, Gauge Blocks
Taper Ring Gauge	M2 to M6 M6 to M20 M20 to M30 M30 to M75 M75 to M100 M100 to M150	0.4 μm 0.5 μm 0.6 μm 2.3 μm 2.4 μm 2.9 μm	Comparison to ULM, Ring Gauges
Taper Thread Plug	M2 to M5 M5 to M11 M11 to M22 M22 to M45 M45 to M180	0.92 μm 1.8 μm 1.1 μm 4.7 μm 7.6 μm	Comparison to ULM, Thread Measuring Wire
¹ Coordinate Measuring Machine X, Y, Z Axis (Linear Accuracy)	Up to 10 mm (10 to 25) mm (25 to 50) mm (50 to 75) mm (75 to 100) mm (100 to 125) mm (125 to 150) mm (150 to 175) mm (175 to 200) mm (200 to 250) mm (250 to 300) mm (300 to 400) mm (400 to 500) mm (500 to 800) mm (800 to 1 000) mm (1 000 to 1 200) mm (1 200 to 1 500) mm	0.16 μm 0.34 μm 0.67 μm 1 μm 1.3 μm 1.7 μm 2.1 μm 2.4 μm 2.7 μm 3.4 μm 4.1 μm 5.4 μm 6.7 μm 11 μm 13 μm 16 μm 20 μm	Comparison to Gauge Blocks

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Ultrasonic Thickness Gauge	Up to 100 mm	6 nm	Comparison to Gauge Blocks
Standard Scale	Up to 50 mm (> 50 to 200) mm (> 200 to 410) mm	1.3 µm 2.4 µm 3.4 µm	Comparison to 3D Vision Measuring Machine, Standard Glass Scale
Coating Thickness Gauge	(30 to 1 470) µm	0.92 µm	Comparison to Calibration Foils
Square	Up to 100 mm (> 100 to 200) mm (> 200 to 300) mm (> 300 to 400) mm (> 400 to 500) mm (> 500 to 700) mm	3.1 µm 3.8 µm 4.5 µm 6 µm 7.2 µm 9.8 µm	Direct Measure using Coordinate Measuring Machine
Spirit Precision Level	10 µm/m 20 µm/m 40 µm/m 50 µm/m 0.1 mm/m	6.8 µm/m 12 µm/m 23 µm/m 29 µm/m 58 µm/m	Comparison to Precision Level Calibrator, Gauge Blocks, Surface Plate
¹ Extensometer	Up to 55 mm	3.6 µm	Comparison to Dial Gauge Tester

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
¹ Electronic Balance, Spring Balance, Load Cell Resolution: 0.000 01 g	Up to 10 g (10 to 20) g (20 to 50) g (50 to 60) g (60 to 70) g (70 to 100) g (100 to 150) g (150 to 220) g (220 to 300) g	0.04 mg 0.05 mg 0.08 mg 0.11 mg 0.12 mg 0.16 mg 0.2 mg 0.3 mg 0.4 mg	OIML Class E2, F1, M1 weight sets and internal calibration procedure utilized in the calibration of the weighing system.

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
¹ Electronic Balance, Spring Balance, Load Cell Resolution: 0.001 g 0.001 g 0.001 g 0.001 g 0.001 g 0.001 g 0.01 g 0.01 g 0.01 g 0.01 kg 0.1 kg 0.5 kg 1 kg 5 kg 10 kg	(300 to 1 000) g (1 000 to 2 000) g (2 to 3) kg (3 to 5) kg (5 to 6) kg (6 to 8) kg (8 to 10) kg (10 to 12) kg (12 to 20) kg (20 to 100) kg (100 to 1 000) kg (1 000 to 5 000) kg (5 000 to 10 000) kg (10 000 to 40 000) kg (40 000 to 80 000) kg	1 mg 2 mg 3 mg 5 mg 10 mg 12 mg 17 mg 20 mg 26 mg 5.8 g 58 g 0.33 kg 0.66 kg 3.2 kg 6.3 kg	OIML Class E2, F1, M1 weight sets and internal calibration procedure utilized in the calibration of the weighing system.
¹ Push-Pull Gauge, Force Gauge, Tension, Tensile	Up to 1 000 N (1 000 to 3 000) N (3 000 to 5 000) N (5 000 to 10 000) N	0.006 N 0.01 N 0.02 N 0.03 N	Comparison to Weight Sets
¹ Hand Torque Tool, Torque Wrench, Torque Driver, Electronic Torque	(0.2 to 20) N·m (20 to 40) N·m (40 to 60) N·m (60 to 80) N·m (80 to 100) N·m (100 to 200) N·m (200 to 400) N·m (400 to 600) N·m (600 to 800) N·m (800 to 1 000) N·m	0.06 N·m 0.07 N·m 0.08 N·m 0.09 N·m 0.1 N·m 3.1 N·m 3.5 N·m 3.9 N·m 4.4 N·m 5 N·m	Comparison to Static Torque Transducer
¹ Hardness Tester, Duro Tester (Types A, B, C, D, DO, O) Indenter Dimensions Length Angle Radius Spring Force	Up to 3.57 mm Up to 36° Up to 0.51 mm Up to 44.5 N	2.1 µm 0.003 4° 3 µm 0.026 N	Direct Verification per ASTM D 2240-15 using Vision Measuring Machine Durometer Calibrator

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Manometer	Up to 1 psig	0.001 9 psi	Comparison to Fluke 700PD2 Pressure Module
¹ Pneumatic Pressure Devices	Up to 200 kPa (> 200 to 2 000) kPa	0.06 kPa 0.58 kPa	Comparison to Pneumatic Pressure Calibrator
¹ Hydraulic Pressure Devices	Up to 7 MPa (> 7 to 70) MPa	4 kPa 21 kPa	Comparison to Hydraulic Pressure Calibrator
¹ Vacuum Gauges	(-90 to 0) kPa	55 Pa	Comparison to Vacuum Calibrator
¹ Barometric Pressure, Absolute Pressure	(100 to 1 034) hPa	0.6 hPa	Comparison to Absolute Pressure Calibrator
Mass (Standard Weights)	1 mg 2 mg 5 mg 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg	6 µg 6 µg 6 µg 13 µg 14 µg 24 µg 26 µg 27 µg 30 µg 12 µg 14 µg 18 µg 24 µg 35 µg 87 µg 0.14 mg 0.26 mg 0.85 mg 2 mg 6.9 mg 11 mg 20 mg 69 mg	Comparison Method using Electronic Balance, OIML Class E2, F1, and M1 Weights
Mass (Standard Weights)	50 kg 100 kg 200 kg 500 kg	1 g 1 g 5 g 5 g	Comparison Method using Electronic Balance, OIML Class F2 Weights

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Torque Calibrators	(0.1 to 2) N·m (2 to 4) N·m (4 to 6) N·m (6 to 8) N·m (8 to 10) N·m (10 to 20) N·m (20 to 40) N·m (40 to 200) N·m (200 to 400) N·m (400 to 1 000) N·m (1 000 to 1 500) N·m	0.92 % of reading 0.46 % of reading 0.31 % of reading 0.23 % of reading 0.18 % of reading 0.1 % of reading 0.05 % of reading 0.03 % of reading 0.02 % of reading 0.01 % of reading 0.006 % of reading	Comparison to Weights, Calibration Arms
¹ Universal Testing Machine, Crane Scales, Compression / Tensile Testing Machine	Compression Testing Machine (100 to 200) kN (> 200 to 1 000) kN Tensile Testing Machine 100 N to 30 kN	0.32 % of reading 0.42 % of reading 0.16 % of reading	Comparison to Master Load Cell
Volumetric Glassware, Burette	5 ml 10 ml 25 ml 50 ml 100 ml	0.003 5 ml 0.003 7 ml 0.006 5 ml 0.01 ml 0.018 ml	Comparison to Electronic Balance
Volumetric Glassware, Volumetric Flask	2 ml 5 ml 10 ml 20 ml 25 ml 50 ml 100 ml 200 ml 250 ml 500 ml 1 000 ml	0.005 8 ml 0.005 8 ml 0.005 9 ml 0.006 2 ml 0.006 5 ml 0.01 ml 0.017 ml 0.028 ml 0.035 ml 0.063 ml 0.13 ml	Comparison to Electronic Balance
Volumetric Glassware, Measuring Cylinder	5 ml 10 ml 25 ml 50 ml 100 ml 250 ml 500 ml 1 000 ml	0.005 8 ml 0.005 9 ml 0.006 5 ml 0.01 ml 0.017 ml 0.035 ml 0.063 ml 0.13 ml	Comparison to Electronic Balance

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Volumetric Glassware, Measuring Pipette	0.5 ml 1 ml 2 ml 5 ml 10 ml 15 ml 25 ml 50 ml	0.002 3 ml 0.002 3 ml 0.002 3 ml 0.002 4 ml 0.003 7 ml 0.006 ml 0.006 5 ml 0.01 ml	Comparison to Electronic Balance
Volumetric Glassware, Volumetric Pipette	0.5 ml 1 ml 2 ml 5 ml 10 ml 15 ml 25 ml 50 ml 100 ml	0.002 3 ml 0.002 3 ml 0.002 3 ml 0.002 4 ml 0.003 7 ml 0.006 ml 0.006 5 ml 0.01 ml 0.016 ml	Comparison to Electronic Balance
Piston Pipette	(10 to 1 000) μ l	0.14 μ l	Comparison to Electronic Balance
¹ Anemometers (Air Velocity)	2.5 m/s 5 m/s 7.5 m/s 10 m/s 12.5 m/s 15 m/s	0.25 m/s 0.25 m/s 0.25 m/s 0.25 m/s 0.42 m/s 0.42 m/s	Comparison to Lutron MHB-382SD Barometer, Trotec TA400 Dynamic Pressure Anemometer, Omega WTM-1000 Wind Tunnel
Liquid Flow Devices	(3 to 6) m ³ /h (6 to 10) m ³ /h (10 to 20) m ³ /h (20 to 30) m ³ /h	0.018 m ³ /h 0.065 m ³ /h 0.069 m ³ /h 0.099 m ³ /h	Comparison to Ultrasonic Flow Meter, Calibration Rig
Liquid Flow Devices	(6 to 12) m ³ /h (12 to 24) m ³ /h (24 to 42) m ³ /h	0.037 m ³ /h 0.056 m ³ /h 0.069 m ³ /h	Comparison to Electromagnetic Flow Meter, Calibration Rig

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Air Flow Meters	(0 to 0.3) lpm	0.006 5 lpm	Comparison to Air Flow Calibrator, Calibration Rig
	(0.3 to 1) lpm	0.012 lpm	
	(1 to 2) lpm	0.021 lpm	
	(2 to 3) lpm	0.031 lpm	
	(3 to 4) lpm	0.04 lpm	
	(4 to 5) lpm	0.05 lpm	
	(5 to 10) lpm	0.1 lpm	
	(10 to 20) lpm	0.2 lpm	
	(20 to 30) lpm	0.3 lpm	
	(30 to 50) lpm	3.4 lpm	
	(50 to 100) lpm	3.8 lpm	
	(100 to 150) lpm	4.1 lpm	
	(150 to 200) lpm	7.5 lpm	
Hydrometers	(600 to 850) kg/m ³	0.08 kg/m ³	Comparison to Electronic Balance, Standard Rong Weight, Barometer
	(> 850 to 1 350) kg/m ³	0.1 kg/m ³	
	(> 1 350 to 2 000) kg/m ³	0.14 kg/m ³	

Photometry and Radiometry

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
^{1,3} Gloss Meters	(0 to 100) GU		Comparison to Standard Gloss Tile
	20°	0.93 GU	
	60°	0.91 GU	
	85°	1.1 GU	
Illuminance/Lux Meter	35 lux	0.46 lux	Comparison to Standard Illuminance/Lux Meter
	50 lux	0.66 lux	
	100 lux	1.3 lux	
	500 lux	6.6 lux	
	1 000 lux	13 lux	
	1 500 lux	20 lux	
	2 000 lux	26 lux	
	3 000 lux	39 lux	
	4 000 lux	52 lux	
	5 000 lux	65 lux	

Photometry and Radiometry

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
⁵ Colorimeter Illuminant A, C, D50, D65 Observer: 2°, 10°	Geometry 8°: DI, 8°: DE		Comparison to Standard Color Tile
	Pale Grey		
	x	0.000 3	
	y	0.000 3	
	Y	0.42	
	u'	0.000 3	
	v'	0.000 3	
	L*	0.21	
	a*	0.14	
	b*	0.14	
	C*	0.18	
	H*	0.18	
	Mid Grey		
	x	0.000 3	
	y	0.000 3	
	Y	0.21	
	u'	0.000 3	
	v'	0.000 3	
	L*	0.21	
	a*	0.14	
	b*	0.18	
	C*	0.21	
	H*	0.22	
	Diff Grey		
	x	0.000 3	
	y	0.000 3	
	Y	0.21	
	u'	0.000 3	
	v'	0.000 3	
	L*	0.21	
	a*	0.14	
	b*	0.18	
	C*	0.21	
	H*	0.22	

Photometry and Radiometry

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
⁵ Colorimeter Illuminant A, C, D50, D65 Observer: 2°, 10°	Geometry 8°: DI, 8°: DE		Comparison to Standard Color Tile
	Deep Grey		
	x	0.000 3	
	y	0.000 3	
	Y	0.14	
	u'	0.000 3	
	v'	0.000 3	
	L*	0.42	
	a*	0.2	
	b*	0.21	
	C*	0.17	
	H*	0.19	
	Deep Pink		
	x	0.001 6	
	y	0.000 6	
	Y	0.21	
	u'	0.001 4	
	v'	0.000 3	
	L*	0.28	
	a*	0.26	
	b*	0.21	
	C*	0.19	
	H*	0.2	
	Red		
	x	0.006 4	
	y	0.000 9	
	Y	0.21	
	u'	0.004 8	
	v'	0.001 4	
	L*	0.42	
	a*	0.35	
	b*	0.71	
	C*	0.68	
	H*	0.92	

Photometry and Radiometry

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
⁵ Colorimeter Illuminant A, C, D50, D65 Observer: 2°, 10°	Geometry 8°: DI, 8°: DE, Orange		Comparison to Standard Color Tile
	x	0.002	
	y	0.000 7	
	Y	0.35	
	u'	0.001	
	v'	0.000 6	
	L*	0.28	
	a*	0.28	
	b*	0.52	
	C*	0.44	
	H*	0.42	
	Bright Yellow		
	x	0.000 7	
	y	0.000 9	
	Y	0.5	
	u'	0.000 3	
	v'	0.000 3	
	L*	0.35	
	a*	0.21	
	b*	0.44	
	C*	0.35	
	H*	0.19	
	Green		
	x	0.000 7	
	y	0.001 4	
	Y	0.21	
	u'	0.000 7	
	v'	0.000 6	
	L*	0.3	
	a*	0.24	
	b*	0.3	
	C*	0.3	
	H*	0.31	

Photometry and Radiometry

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
⁵ Colorimeter Illuminant A, C, D50, D65 Observer: 2°, 10°	Geometry 8°: DI, 8°: DE, Diff. Green		Comparison to Standard Color Tile
	x	0.000 7	
	y	0.001 4	
	Y	0.21	
	u'	0.000 7	
	v'	0.000 6	
	L*	0.3	
	a*	0.24	
	b*	0.3	
	C*	0.3	
	H*	0.31	
	Cyan		
	x	0.001 7	
	y	0.000 7	
	Y	0.28	
	u'	0.001	
	v'	0.000 6	
	L*	0.28	
	a*	0.28	
	b*	0.21	
	C*	0.28	
	H*	0.28	
	Deep Blue		
	x	0.026	
	y	0.029	
	Y	0.14	
	u'	0.008 5	
	v'	0.028	
	L*	1.2	
	a*	1.8	
	b*	1.4	
	C*	2.9	
	H*	2.5	

Photometry and Radiometry

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
⁵ Colorimeter Illuminant A, C, D50, D65 Observer: 2°, 10°	Geometry 0°:45°, Pale Grey		Comparison to Standard Color Tile
	x	0.000 3	
	y	0.000 3	
	Y	0.71	
	u'	0.000 3	
	v'	0.000 3	
	L*	0.42	
	a*	0.14	
	b*	0.14	
	C*	0.17	
	H*	0.18	
	Mid Grey		
	x	0.000 3	
	y	0.000 3	
	Y	0.35	
	u'	0.000 3	
	v'	0.000 3	
	L*	0.35	
	a*	0.14	
	b*	0.14	
	C*	0.2	
	H*	0.2	
	Diff. Grey		
	x	0.000 3	
	y	0.000 3	
	Y	0.35	
	u'	0.000 3	
	v'	0.000 3	
	L*	0.35	
	a*	0.14	
	b*	0.14	
	C*	0.2	
	H*	0.2	

Photometry and Radiometry

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
⁵ Colorimeter Illuminant A, C, D50, D65 Observer: 2°, 10°	Geometry 0°:45°, Deep Grey		Comparison to Standard Color Tile
	x	0.000 5	
	y	0.000 3	
	Y	0.28	
	u'	0.000 3	
	v'	0.000 3	
	L*	0.71	
	a*	0.14	
	b*	0.2	
	C*	0.19	
	H*	0.2	
	Deep Pink		
	x	0.001 6	
	y	0.000 6	
	Y	0.28	
	u'	0.001 4	
	v'	0.000 3	
	L*	0.42	
	a*	0.28	
	b*	0.21	
	C*	0.17	
	H*	0.19	
	Red		
	x	0.006 4	
	y	0.000 9	
	Y	0.35	
	u'	0.004 8	
	v'	0.001 4	
	L*	0.57	
	a*	0.42	
	b*	1.3	
	C*	0.86	
	H*	1	

Photometry and Radiometry

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
⁵ Colorimeter Illuminant A, C, D50, D65 Observer: 2°, 10°	Geometry 0°:45°, Orange		Comparison to Standard Color Tile
	x	0.002	
	y	0.000 7	
	Y	0.57	
	u'	0.001	
	v'	0.000 6	
	L*	0.42	
	a*	0.28	
	b*	0.86	
	C*	0.44	
	H*	0.41	
	Bright Yellow		
	x	0.000 7	
	y	0.000 8	
	Y	0.78	
	u'	0.000 3	
	v'	0.000 3	
	L*	0.42	
	a*	0.21	
	b*	0.42	
	C*	0.32	
	H*	0.2	
	Green		
	x	0.000 7	
	y	0.001 4	
	Y	0.28	
	u'	0.000 7	
	v'	0.000 6	
	L*	0.35	
	a*	0.21	
	b*	0.28	
	C*	0.29	
	H*	0.3	

Photometry and Radiometry

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
⁵ Colorimeter Illuminant A, C, D50, D65 Observer: 2°, 10°	Geometry 0°:45°, Diff. Green		Comparison to Standard Color Tile
	x	0.000 7	
	y	0.001 4	
	Y	0.28	
	u'	0.000 7	
	v'	0.000 6	
	L*	0.35	
	a*	0.21	
	b*	0.28	
	C*	0.29	
	H*	0.3	
	Cyan		
	x	0.001 7	
	y	0.000 7	
	Y	0.35	
	u'	0.001	
	v'	0.000 6	
	L*	0.42	
	a*	0.28	
	b*	0.28	
	C*	0.32	
	H*	0.32	
	Deep Blue		
	x	0.026	
	y	0.029	
	Y	0.28	
	u'	0.008 6	
	v'	0.028	
	L*	2.1	
	a*	3.2	
	b*	2.5	
	C*	3	
	H*	3.2	

Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
¹ Temperature Controlled Chamber, Hot Air Oven, Incubator, Refrigerator, Low Temperature Incubator, Autoclave	(-40.0 to 0) °C (0 to 100) °C (100 to 200) °C (200 to 250) °C	0.27 °C 0.19 °C 0.26 °C 0.31 °C	Comparison to Agilent 34970A Data Logger, Thermocouple, RTD sensor
¹ Temperature Gauge, Dial Thermometer	(-80 to 400) °C	0.07 °C	Comparison to Hart Scientific 1575 PRT Standard
¹ Temperature Gauge, Dial Thermometer	(400 to 650) °C	2.6 °C	Comparison to Fluke 1524 Thermocouple Standard
¹ Thermocouple Sensor TC	(-80 to 400) °C	0.07 °C	Comparison to Hart Scientific 1575 PRT Standard, Fluke 744 Documenting Process Calibrator
¹ Thermocouple Sensor TC	(400 to 1 200) °C	2.6 °C	Comparison to Fluke 1524 Thermocouple Standard, Fluke 744 Documenting Process Calibrator
¹ Liquid Bath	(-40.0 to 0) °C (0 to 100) °C (100 to 200) °C (200 to 250) °C	0.27 °C 0.19 °C 0.26 °C 0.31 °C	Comparison to Agilent 34970A Data logger with RTD sensor
¹ Digital Thermometer with Thermocouple Sensors Types K, J, E, T, N, R, S	(-80 to 400) °C	0.07 °C	Comparison to Hart Scientific 1575 PRT Standard
¹ Digital Thermometer with Thermocouple Sensors Types K, J, E, T, N, R, S	(400 to 1 200) °C	2.6 °C	Comparison to Fluke 1524 Thermocouple Standard
¹ Digital Thermometer with RTD or Thermistor Sensor	(-80 to 400) °C	0.07 °C	Comparison to Hart Scientific 1575 PRT Standard
¹ Digital Thermometer with RTD or Thermistor Sensor	(400 to 850) °C	2.6 °C	Comparison to Fluke 1524 Thermocouple Standard

Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
¹ RTD Sensors	(-80 to 400) °C	0.07 °C	Comparison to Hart Scientific 1575 PRT Standard, Fluke 744 Documenting Process Calibrator
¹ RTD Sensors	(400 to 850) °C	2.6 °C	Comparison to Fluke 1524 Thermocouple Standard, Fluke 744 Documenting Process Calibrator
Liquid-in-Glass Thermometers	(-80 to 250) °C	0.29 °C	Comparison to Hart Scientific 1575 PRT Standard
Dry Block, Dry Well	Up to 250 °C	0.07 °C	Comparison to Hart Scientific 1575 PRT Standard, Fluke 1524 Thermocouple Standard
Dry Block, Dry Well	(250 to 450) °C (450 to 1 200) °C	0.7 °C 2.6 °C	Comparison to Hart Scientific 1575 PRT Standard, Fluke 1524 Thermocouple Standard
¹ Furnace	(300 to 600) °C (600 to 900) °C (900 to 1 200) °C	1.8 °C 2.5 °C 2.6 °C	Comparison to Thermocouple Standard
Digital Thermometer with Surface Probe	(40 to 350) °C	2.4 °C	Comparison to Digital Thermometer Fluke 714 with Surface Probe
Infrared Thermometers	(-40 to 50) °C (> 50 to 100) °C (> 100 to 200) °C (> 200 to 400) °C	0.91 °C 0.92 °C 1.5 °C 2 °C	Comparison to In-House Black Body in Liquid Bath $\epsilon = 0.95$, $\lambda = (8 \text{ to } 14) \mu\text{m}$
Thermo Hygrometer Temperature	(15 to 40) °C	0.2 °C	Comparison to Fluke 5020A Thermo-hygrometer, Temp/Humidity Chamber

Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Thermo Hygrometer Humidity	(30 to 50) %RH (50 to 70) %RH (70 to 90) %RH	0.84 %RH 1.1 %RH 1.6 %RH	Comparison to Fluke 5020A Thermo-hygrometer, Temp/Humidity Chamber
¹ Thermo Hygrometer Temperature Chamber	(20 to 40) °C	0.11 °C	Comparison to Agilent 34970A Datalogger with RTD Sensor
¹ Thermo Hygrometer Humidity Chamber	(30 to 70) %RH	3.3 %RH	Comparison to Data Logger CEM DT-172

Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
^{1,3} Digital Photo Tachometers	(2.5 to 999.9) rpm (> 999.9 to 9 999.9) rpm (> 9 999.9 to 99 999) rpm	0.01 rpm 0.06 rpm 0.58 rpm	Comparison to Fluke 5502A Multiproduct Calibrator with LED
^{1,3} Digital Contact Tachometers	(0.5 to 999.9) rpm (> 999.9 to 9 999.9) rpm (> 9 999.9 to 19 999) rpm	0.01 rpm 0.06 rpm 0.58 rpm	Comparison to Fluke 5502A Multiproduct Calibrator
^{1,3} Stroboscopes	Up to 120 rpm (> 120 to 1 020) rpm (> 1 020 to 5 040) rpm (> 5 040 to 10 020) rpm (> 10 020 to 20 040) rpm (> 20 040 to 50 040) rpm (> 50 040 to 99 960) rpm	0.1 rpm 0.1 rpm 0.22 rpm 0.36 rpm 0.65 rpm 1.5 rpm 3 rpm	Comparison to Fluke PM6685 Universal Counter, High-Speed Photo Transistor Box (PT523C-B1-T363)
¹ Stopwatch Photo Totalize Method Time Base Method	10 s to 1 h (1 to 86 400) s	27 ms 0.58 ms	Comparison to Agilent 53132A Universal Frequency Counter, Fluke PM6685R Universal Frequency Counter, HP 8904A Multifunction Synthesizer
¹ Frequency Source	(0.01 to 500) Hz 500 Hz to 5 kHz (5 to 50) kHz	20 µHz/Hz + 5.9 mHz 20 µHz/Hz + 58 mHz 20 µHz/Hz + 0.58 Hz	Comparison to Fluke 5502A Multiproduct Calibrator

Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
³ General Frequency Source	(1 to 1 000) Hz > 1 Hz to 10 kHz (> 0.01 to 225) MHz (> 225 to 300) MHz > 300 MHz to 1 GHz (> 1 to 1.8) GHz (> 1.8 to 18) GHz	$4.4 \times 10^{-10} f$ $2.4 \times 10^{-10} f$ $2.4 \times 10^{-10} f$ $2.4 \times 10^{-9} f$ $7 \times 10^{-10} f$ $4.2 \times 10^{-10} f$ $2.7 \times 10^{-10} f$	Comparison to Agilent 53132A Universal Frequency Counter, Fluke PM6685R Universal Frequency Counter, HP 8902A Measuring Receiver
Radar Gun Speed	24.150 GHz 40.25 km/h 56.35 km/h 104.65 km/h 34.7 GHz 40.64 km/h 64.75 km/h	1 km/h 1 km/h	Comparison to Tuning Forks
¹ Radar Speed (All Frequency Band)	60 km/h 90 km/h 120 km/h	0.4 km/h 0.6 km/h 0.9 km/h	Rasmi Racing Drag and Timer, Measuring Tape, Calculate speed (S) by known distance (D) and known elapse time (T) $S = \frac{D}{T}$
^{1,3} Centrifuge Rotation	(50 to 999.99) rpm (1 000 to 3 000) rpm (3 000.1 to 9 999.9) rpm (10 000 to 20 000) rpm	0.53 rpm 1.9 rpm 5.3 rpm 13 rpm	Comparison to Digital Tachometer

Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Universal Frequency Counter Time Base	1 MHz to 10 MHz	2.4×10^{-10} Hz	Comparison to Agilent 53132A Universal Frequency Counter, Fluke PM6685R Universal Frequency Counter,
Frequency	DC to 18 GHz	2.9×10^{-10} Hz	Agilent N9310A RF Signal Generator, HP 33120A Function
Time Interval	1 μ s to 1 ms	0.6 ps	Generator/Arbitrary Waveform Generator, HP 83731A Synthesizer Signal
Trigger Level	(-5.25 to 5.25) V	1.9 μ V	Generator, HP 8904A Multifunction Synthesizer
Time Interval Source	10 ns to 1 s (> 1 to 10) s (> 10 to 50) s (> 50 to 100) s	2 ns 4 ns 14 ns 26 ns	Comparison to Agilent 53132A Universal Frequency Counter, Fluke PM6685R Universal Frequency Counter

DIMENSIONAL MEASUREMENT

1 Dimensional

Parameter	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Jig, Fixture and Mold, Die X-axis	Up to 25 mm (25 to 50) mm (50 to 75) mm (75 to 100) mm (100 to 125) mm (125 to 150) mm	2.7 μ m 2.8 μ m 2.9 μ m 3 μ m 3.2 μ m 3.4 μ m	Direct 1D Measurements using a Coordinate Measuring Machine

1 Dimensional


Parameter	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Jig, Fixture and Mold, Die X-axis	(150 to 175) mm (175 to 200) mm (200 to 250) mm (250 to 300) mm (300 to 400) mm	3.6 μ m 3.8 μ m 4.3 μ m 4.9 μ m 6 μ m	Direct 1D Measurements using a Coordinate Measuring Machine
Jig, Fixture and Mold, Die X-axis	(400 to 500) mm (500 to 650) mm	7.2 μ m 8.5 μ m	Direct 1D Measurements using a Coordinate Measuring Machine
Jig, Fixture and Mold, Die Y-axis	Up to 25 mm (25 to 50) mm (50 to 75) mm (75 to 100) mm (100 to 125) mm (125 to 150) mm (150 to 175) mm (175 to 200) mm (200 to 250) mm (250 to 300) mm (300 to 400) mm (400 to 500) mm (500 to 650) mm	2.7 μ m 2.8 μ m 2.9 μ m 3 μ m 3.2 μ m 3.4 μ m 3.6 μ m 3.8 μ m 4.3 μ m 4.9 μ m 6 μ m 7.2 μ m 9.8 μ m	Direct 1D Measurements using a Coordinate Measuring Machine
Jig, Fixture and Mold, Die Z-axis	Up to 25 mm (25 to 50) mm (50 to 75) mm (75 to 100) mm (100 to 125) mm (125 to 150) mm (150 to 175) mm (175 to 200) mm (200 to 250) mm (250 to 300) mm (300 to 400) mm (400 to 500) mm	2.7 μ m 2.8 μ m 2.9 μ m 3 μ m 3.2 μ m 3.4 μ m 3.6 μ m 3.8 μ m 4.3 μ m 4.9 μ m 6 μ m 7.2 μ m	Direct 1D Measurements using a Coordinate Measuring Machine

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

- On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
- Nominal Values listed as approximate.
- f = frequency in Hz; " = arc-second; GU = Gloss Units; rpm = revolutions per minute.
- Mismatch Uncertainty is based on DUT SWR: 1.4 for < 2 GHz; 1.6 for < 18 GHz.
- Unitless parameter.
- Unless otherwise specified in the far-right column, the calibration procedure utilized in the calibration of the parameter was developed internally.

7. This scope is formatted as part of a single document including Certificate of Accreditation No. ACT-2050.



Jason Stine, Vice President

